

Terms and Conditions :

1. The item which is not matching with specifications will not be accepted.
2. Sub-standard/defective piece will not be accepted at any cost. Such piece is to be replaced at Firms own risk and cost.
3. The firm shall enclose an affidavit that it has not been black listed by Central/State Government/any PSUs/Private Sector.
4. The interested company may submit its tender along with Earnest Money o f Rs. 10,000/- (Rupees Ten thousand only) in the form of DD/Pay Order drawn in favour of Assistant Director ,Regional Ayurveda Research Institute for Urinary Disorders, Bantalab, Jammu.
5. Tenderers will have to furnish a copy of their PAN No., TIN No. and Sale Tax Registration Number (proof to be attached).
6. Purchaser will not pay any amount separately for transit insurance till the entire stores contracted for, arrive in good condition at the destination. The supplier will be responsible for demo and installation of the instrument in the premises of RARIUD.

The Assistant Director reserves the right to accept or reject quotation/quotations without assigning any reason thereof.

Part of Technical Bid (Detailed Technical Specifications of the equipments.)

Compliance Chart (to be submitted by the tenderer)

	Detailed Specifications				Compliance in (Yes or No only)
					Yes/No
Optical system	UIS2 (Universal Infinity corrected) optical system				
Illumination	Built-in-transmitted Koehler illuminator 6V30W halogen bulb 100-120V/220-240V~ 0.85/0.45A 50/60 Hz				
Focusing	<ul style="list-style-type: none"> • Stage height movement by roller guide (rack & pinion) • Stroke per rotation: 36.8 mm • Full stroke range : 25 mm • Upper limit stopper • Tension adjustment on coarse focus adjustment knob 				
Revolving nose piece	Fixed quintuple nosepiece with inward tilt				
Observation tube	Type	U-CBI30-2, Binocular	U-CTR30-2, Trinocular	U-CTBI, Tilting Binocular	
	Field number	20	20	18	
	Tube inclination	30 ⁰	30 ⁰	30 ⁰ -60 ⁰	
	Inter pupillary distance adjustment range	48-75 mm	48-75 mm	48-75 mm	

	Light path selector	None	None (Bi 50%, Video/Photo 50% fixed)	None	
Stage	Size	188 (W) X 134 (D) mm			
	Movement range	76 mm X-Direction X 50 mm Y-Direction			
	Specimen holder	Double slide holder			
	Rubber grip	Equipped as standard			
Condenser (CH3-CD)	Type	Abbe condenser			
	N.A.	1.25 with oil immersion			
	Aperture iris diaphragm	Built-in			
Dimensions & Weight	233 (W) X 432 (H) X 367.5 (D) mm, approximately 8.5 Kg (appr 18.7 lb)				
AC adaptor	Input Voltage 220/110 AC Output Voltage 7.5V DC / 12V DC Power Max.15-18W				

CCD Camera attachment	<p>Must require Trinocular Head and adaptor as per microscope model</p> <p>Image stitching</p> <p>Extended depth of focus</p> <p>Capture images, videos, and time elapse images</p> <p>Compatible with Windows, Mac Os and Linux</p>		
Desktop	HP desktop with approx 19 inch LED Screen	Compatible for Trinocular version microscope.	

PROFORMA OF FINANCIAL BID

(In sealed Cover-II super scribed "Financial Bid")

(To be typed on Letter Head of the Tenderer, duly stamped and signed)

Our quoted rate for supplying Item at RARIUD , Rajinder Nagar Ban Talab Jammu will be as follows.

Name of Equipment	Basic Rates (with warranty)	Taxes	Total Landing cost FOR Stores	Total Rate